
**Medical devices — Symbols to be used
with medical device labels, labelling and
information to be supplied —**

**Part 1:
General requirements**

AMENDMENT 1

*Dispositifs médicaux — Symboles à utiliser avec les étiquettes,
l'étiquetage et les informations à fournir relatifs aux dispositifs
médicaux —*

Partie 1: Exigences générales

AMENDEMENT 1



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2008

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

Amendment 1 to ISO 15223-1:2007 was prepared by Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*.

STANDARDSISO.COM : Click to view the full PDF of ISO 15223-1:2007/Amd 1:2008





Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied —




Part 1: General requirements

AMENDMENT 1

Page 8, Table 1

Add the following symbols to Table 1.

No.	Symbol	Title	ISO 7000 or IEC 60417 registration number
5.32		Sampling site	ISO 7000-2715
5.33		Fluid path	ISO 7000-2722
5.34		Non-pyrogenic	ISO 7000-2724
5.35		Contains or presence of natural rubber latex NOTE This symbol should be used only when natural rubber latex is a material of construction within the device or the packaging of a device. It is intended to warn those people who may have allergic reactions to certain proteins in natural rubber latex. This symbol should not be used for devices containing "synthetic" rubber.	Derived from ISO 7000-2725

No.	Symbol	Title	ISO 7000 or IEC 60417 registration number
5.36		<p>Drops per millilitre</p> <p>NOTE On medical devices: to indicate the number of drops per millilitre. The number of drops per millilitre is specified; "20" is shown as an example and should be replaced by the correct number of drops per millilitre.</p>	ISO 7000-2726
5.37		<p>Liquid filter with pore size</p> <p>NOTE On medical devices: to indicate that the infusion or transfusion system contains a liquid filter in various sizes. The nominal pore size of the filter is specified; "15" is shown as an example and should be replaced by the correct pore size.</p>	ISO 7000-2727
5.38		<p>One-way valve</p>	ISO 7000-2728