
**Cardiovascular implants and artificial
organs — Hard-shell cardiectomy/
venous reservoir systems (with/
without filter) and soft venous
reservoir bags**

AMENDMENT 1: Connectors

*Implants cardiovasculaires et organes artificiels — Systèmes
réservoirs de cardiectomie/veineux à paroi dure (avec/sans filtre) et
sacs réservoirs veineux mous*

AMENDEMENT 1: Raccords





COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

STANDARDSISO.COM : Click to view the full PDF of ISO 15674:2016/Amd 1:2020

Cardiovascular implants and artificial organs — Hard-shell cardiectomy/venous reservoir systems (with/without filter) and soft venous reservoir bags

AMENDMENT 1: Connectors

4.2.3 Connectors

Replace the text of 4.2.3 with the following text:

Connectors for connection to the blood pathway shall, when tested in accordance with 5.3.3, allow a secure connection (see [Figures B.1](#) through [B.11](#) for examples of connectors).

NOTE 1 Connectors of a type that allows connection of tubes with an inner diameter of 4,8 mm, 6,3 mm, 9,5 mm or 12,7 mm, a type that complies with ISO 8637-1:2017, Figure 1, or a type that complies with ISO 80369-7 have been found satisfactory.

NOTE 2 Connectors with dimensions as given in [Annex B](#) and fitting to functional gauges and reference steel fittings is a way to comply with this requirement.

Performance testing of the connectors shall be performed according to ISO 80369-7: 2016, Clause 6, using the reference fittings given in [Annex B](#).

Normative References

Add:

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

Annex B


Add the following annex after Annex A, before the Bibliography:

Annex B (informative)


Examples of connectors and reference fittings

B.1 Luer Slip Fittings

B.1.1 Figures B.1 and B.2 depict Luer slip fittings. For corresponding dimensions, see Table B.1.




a) Male 6 % (Luer) conical fitting ("male fitting")



b) Female 6 % (Luer) conical fitting ("female fitting")

NOTE See Key and dimensions given in Table B.1.

Figure B.1 — Typical 6 % (Luer) conical fittings



NOTE See Key and dimensions given in Table B.1.

Figure B.2 — Typical assembly of 6 % (Luer) conical fittings

Table B.1 — Dimensions of 6 % (Luer) conical fittings

Reference			Designation	Dimension (length in mm)	
				Rigid material	Semi-rigid material
A			Male fitting	N/A	N/A
B			Female fitting	N/A	N/A
Basic dimensions	<i>d</i>	min.	Minimum diameter of the end of the male conical fitting (reference diameter)	3,925	3,925
		max.	Maximum diameter at the end of the male conical fitting	3,990	4,027
	<i>D</i>	min.	Minimum diameter at the opening of the female conical fitting	4,270	4,270
		max.	Maximum diameter at the opening of the female conical fitting	4,315	4,315
	<i>E</i>		Minimum length of the male conical fitting	7,500	7,500
	<i>F</i>		Minimum depth of the female conical fitting	7,500	7,500
	<i>G</i>		Maximum outside diameter of female conical fitting	6,730	6,730
Other dimensions	<i>L</i> ^a		Minimum length of engagement	4,665	4,050
	<i>M</i> ^a		Tolerance for length of engagement of the female conical fitting	0,750	0,750
	<i>N</i> ^a		Tolerance for length of engagement of the male conical fitting	1,083	1,700
	<i>R</i> ^b		Radius of curvature (maximum)	0,5	0,5
^a Dimensions <i>L</i> , <i>M</i> and <i>N</i> are derived from the basic dimensions.					
^b Or equivalent entry chamfer without any sharp corners.					

B.1.2 Gauging test

B.1.2.1 When tested in accordance with [B.1.2.4](#), the conical fitting should satisfy the requirements specified in [B.1.2.2](#) and [B.1.2.3](#).

B.1.2.2 The small end of the male conical fitting should lie between the two limit planes of the gauge and the larger end of the tapered portion should extend beyond the datum plane of the gauge. Rocking should not be evident between the gauge and the fitting made of rigid material undergoing test.

B.1.2.3 The plane of the maximum diameter at the opening of the female conical fitting should lie between the two limit planes of the gauge. Rocking should not be evident between the gauge and the fitting made of rigid material undergoing test.


B.1.2.4 The procedure should be carried out as specified in [B.1.2.4.1](#) to [B.1.2.4.4](#).

B.1.2.4.1 Carry out the test using steel gauges as illustrated in [Figure B.3](#).

B.1.2.4.2 Carry out the test at a temperature of $(20 \pm 5) ^\circ\text{C}$.

B.1.2.4.3 Prior to testing, condition products made from hygroscopic materials at $(20 \pm 5) ^\circ\text{C}$ and $(50 \pm 10) \%$ relative humidity for not less than 24 h. Conditioning is not required for products made from non-hygroscopic materials.

B.1.2.4.4 Apply the gauge to the conical fitting with a total axial force of 5 N, without the use of torque. Remove the axial load.



a) Gauge for testing rigid male conical fittings

b) Gauge for testing semi-rigid male conical fittings


c) Gauge for testing female conical fittings of all materials

NOTE Cone taper (0,06:1).

Figure B.3 — Gauges for testing 6 % (Luer) conical fittings


B.1.3 Reference steel fittings

B.1.3.1 Figures B.4 and B.5 depict male and female reference steel fittings.



NOTE Cone taper (0,06:1).

Figure B.4 — Reference steel female conical fitting



NOTE Cone taper (0,06:1).


Figure B.5 — Reference steel male conical fitting

B.2 Luer Lock Fittings

B.2.1 Figures B.6 through B.9 depict Luer lock fittings while Figures B.10 and B.11 depict female reference steel fittings for testing male 6 % (Luer) lock fittings. For corresponding dimensions see Table B.2.

If a female 6 % (Luer) conical lock fitting as shown in Figure B.8 has lugs in a plane inclined to the axis of fitting, the lugs should form a part of the thread form shown in Figure B.9. In this case, 'V' does not apply.

All outside edges of lug or thread form as shown in Figures B.10 and B.11 should have a radius between 0,15 mm and 0,2 mm (unless otherwise specified).




Key

1 double start, right-hand thread of 2,5 mm pitch

NOTE See Key and dimensions given in Table B.2.


Figure B.6 — Male 6 % (Luer) conical lock fitting with permanently connected internally threaded collar





NOTE 1 For other dimensions, see Figure B.6.

NOTE 2 See Key and dimensions given in Table B.2.


Figure B.7 — Male 6 % (Luer) conical lock fitting with rotatable internally threaded collar



a) Variant A



b) Variant B




c) Variant C

NOTE 1 Variants B and C are intended to be used for the design of rigid fittings only.

NOTE 2 To ensure compatibility with existing rigid fittings, a maximum $K = 0,8$ mm is preferred.

NOTE 3 See Key and dimensions given in [Table B.2](#).

Figure B.8 — Female 6 % (Luer) conical lock fittings with lugs in a plane at right angles to axis of fitting



Key

1 pitch

NOTE 1 For other dimensions, see [Figure B.8](#).

NOTE 2 See Key and dimensions given in [Table B.2](#).

Figure B.9 — Female 6 % (Luer) lock conical fitting with external thread

Table B.2 — Dimensions of 6 % (Luer) rigid conical lock fittings

Dimensions in millimetres

Symbol	Designation	Dimensions	
		Figures B.6, B.7, B.8 a) and B.9	Figures B.8 b) and B.8 c)
α	Angle of thread or lug bearing surface against separation with the plane perpendicular to the axis of lock fitting	$25^{\circ+5^{\circ}}_0$	$25^{\circ+5^{\circ}}_0$
β	Minimum angle of internal thread non-bearing surface against separation with the plane perpendicular to the axis of lock fitting	25°	—
γ	Minimum angle of external thread or lug non-bearing surface against separation with the plane perpendicular to the axis of the lock fitting	0°	0°
E	Minimum length of male lock fitting	7,5	—
F	Nominal distance from the face of the fitting to the base of the lug	—	0,20
G	Maximum outside diameter of female lock fitting at base of lugs or maximum inside diameter of external thread; this diameter should not be increased for a distance from the hub face of 5,5 mm	6,73	5,7
H	Root diameter of the thread of male lock fitting	$8,0 \pm 0,1$	—
J	Crest diameter of the thread of male lock fitting	$7,0 \pm 0,2$	—
K	Maximum thread width of male lock fitting at root	1	—
P	Minimum projection of nozzle from collar	2,1	—

Table B.2 (continued)

Symbol	Designation	Dimensions	
		Figures B.6, B.7, B.8 a) and B.9	Figures B.8 b) and B.8 c)
<i>Q</i>	Minimum thread crest width of male lock fittings	0,3	—
<i>S</i>	Lug crest width or thread crest width of female lock fitting with lugs or external thread	0,3 min.	0,27 max.
<i>T</i>	Maximum distance from tip of male lock fitting to the bottom of first complete thread form of the internal thread	3,2	—
<i>V</i>	Maximum chord length at base of lug in a plane at right angles to axis of fitting only, to be measured on a chord of a circle the diameter of which is <i>J</i> min. (7,0 mm)	3,5	5,0
<i>W</i>	Minimum chord length at extremity of lug in a plane at right angles to axis of fitting only (<i>W</i> should not be greater than <i>V</i>)	2,71	—
<i>X</i>	Distance from axis of female lock fitting to extremity of lug	—	—
<i>2X</i>	Outside diameter across the lugs or external thread	$7,83_{-0,1}^0$	$7,80_{-0,1}^0$
<i>Y</i>	Maximum width of base of lug (axial) or thread at base, of female lock fitting to be measured at a point corresponding to an outside diameter equal to <i>G</i> (6,73 max.)	1,2	1,30
<i>Z</i>	Width across the lugs at external thread	—	$6,50_{-0,1}^0$
Pitch	Nominal pitch of double-start, right-hand thread of female lock fitting — 5 mm lead	2,5	—