

INTERNATIONAL STANDARD

ISO
8454

Second edition
1995-11-15

Cigarettes — Determination of carbon monoxide in the vapour phase of cigarette smoke — NDIR method

*Cigarettes — Dosage du monoxyde de carbone dans la phase gazeuse de
fumée de cigarette — Méthode IRND*



Reference number
ISO 8454:1995(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 8454 was prepared by Technical Committee ISO/TC 126, *Tobacco and tobacco products*.

This second edition cancels and replaces the first edition (ISO 8454:1987), which has been technically revised.

© ISO 1995

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Cigarettes — Determination of carbon monoxide in the vapour phase of cigarette smoke — NDIR method

1 Scope

This International Standard specifies a method for the determination of carbon monoxide (CO) in the vapour phase of cigarette smoke.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3308:1991, *Routine analytical cigarette-smoking machine — Definitions and standard conditions*.

ISO 3402:1991, *Tobacco and tobacco products — Atmosphere for conditioning and testing*.

ISO 4387:1991, *Cigarettes — Determination of total and nicotine-free dry particulate matter using a routine analytical smoking machine*.

ISO 8243:1991, *Cigarettes — Sampling*.

3 Definitions

For the purposes of this International Standard, the following definitions apply.

3.1 vapour phase: The portion of smoke which passes the particulate phase trap during smoking in accordance with ISO 4387 using a machine conforming to ISO 3308.

3.2 clearing puff: Any puff taken after a cigarette has been extinguished or removed from the cigarette holder.

4 Principle

Smoking of cigarettes in accordance with the procedure given in ISO 4387. Collection of the vapour phase of the cigarette smoke and measurement of the carbon monoxide using a non-dispersive infrared (NDIR) analyser calibrated for carbon monoxide. Calculation of the amount of carbon monoxide per cigarette.

5 Apparatus

Usual laboratory apparatus and, in particular, the following items.

5.1 Conditioning enclosure, maintained accurately in accordance with the conditions specified in ISO 3402, for conditioning the cigarette sample prior to smoking (see also 8.1).

5.2 Routine analytical cigarette-smoking machine and accessories, complying with the requirements of ISO 3308.

5.3 Vapour-phase collection system, which can be fitted to one or more of the smoking machine channels. The use of the system shall ensure collection of all the vapour phase (normally vented to atmosphere) to be stored in a previously evacuated container for subsequent sampling through an NDIR analyser.

The collection system shall not cause interference with the normal performance of the smoking machine

and the consequent determination of total particulate matter and nicotine.

The impermeability of the gas-collecting device to a vapour phase shall be checked with a vapour phase containing 4 % to 6 % (V/V) of CO. The CO concentration shall be measured directly after filling the previously evacuated gas-collecting device. After a period of not less than 2 h, the measured value of CO concentration of the vapour phase in the device shall not differ by more than 0,2 % from the value expected from the first determination.

When a bag is used as gas-collecting device, it should be large enough to avoid the final pressure of its contents exceeding the ambient atmospheric pressure.

NOTE 1 The volume of the bag should also be not greater than twice the volume of the gas content collected at atmospheric pressure. In practice, the collection of the vapour phase from 5 cigarettes requires a volume of 3 litres and the collection of the vapour phase from 20 cigarettes requires a volume of 10 litres.

5.4 Non-dispersive infrared (NDIR) analyser, selective and calibrated for the measurement of carbon monoxide in vapours and gases.

Analysers are available from several manufacturers and should have a preferred working range of 0 % to 10 % (V/V) CO and a sampling rate of between 0,5 l/min and 2 l/min. The analyser shall have a precision of 1 % of full scale, a linearity of 1% of full scale and a repeatability of 0,2 % of full scale, under conditions of constant temperature and pressure. Its response to 10 % (V/V) of CO₂ shall not exceed 0,05 % (V/V) as CO. Its response to 2 % (V/V) of water vapour shall not exceed 0,05 % (V/V) as CO.

5.5 Ignition device

Flameless ignition shall be used. The lighters shall light the cigarettes at the first attempt without either touching or pre-charring the cigarettes.

5.6 Barometer, capable of measuring atmospheric pressures to the nearest 0,1 kPa.

5.7 Thermometer, capable of measuring temperature to the nearest 0,2 °C.

6 Standard gas mixtures

Make-up gases other than nitrogen, such as helium, shall not be used since the detected response of carbon monoxide would be different.

NOTE 2 Gases used should be of high purity and used within the manufacturer's time limits.

The NDIR analyser should be calibrated with at least three standard gas mixtures of accurately known

concentrations within a relative error of 2 % covering the expected range in such a way as to avoid extrapolation of the calibration curve; typically about 1 %, 3 % and 5 % (V/V) of CO in nitrogen would be appropriate.

7 Sampling

Sampling shall be carried out in accordance with ISO 8243.

8 Procedure

8.1 Conditioning

Condition the test portion taken from and representative of the laboratory sample in accordance with ISO 3402. Verify that equilibrium has been properly attained as described in ISO 3402.

The atmosphere in the laboratory where the smoking is to be carried out shall also be in accordance with ISO 3402. Place the conditioned test portion in an airtight container (just large enough to contain the portion) and remove each cigarette from the container just before smoking.

8.2 Calibration of the NDIR analyser

8.2.1 Warm up the instrument according to the manufacturer's recommendations, purge the instrument with air and adjust to read zero.

8.2.2 Fill a previously evacuated vapour-phase collection container with the standard gas mixture of nominally 5 % (V/V) CO, re-evacuate and refill with gas. Ensure that the gas in the container is at ambient temperature and pressure. Introduce the gas into the measuring cell using the system sampling pump allowing 5 s to 10 s for equilibration of pressure of the analyser. Note the reading on the analyser concentration indicator when a steady value has been obtained.

If necessary, adjust the instrument reading to agree with the certified value of the standard gas.

8.2.3 Repeat the procedure as specified in 8.2.2 for at least two other standard gas mixtures. If there is a difference of greater than 0,2 % (V/V) CO between the observed and expected values, attention should be given to the analyser linearity.

8.2.4 Recalibrate the instrument at least once a week, using the standard gases. The calibration shall be linear within the limits reported in 5.4.

8.2.5 Check the calibration prior to the measurement using the standard gas containing about 5 % (V/V) carbon monoxide. If there is a difference of greater than 0,2 % (V/V) CO between observed and expected values, repeat the full calibration (8.2).

8.3 Smoking and collection of vapour phase

8.3.1 Preparation of vapour-phase collection system

Prepare the system using the instructions pertinent to the equipment fitted.

Ensure that the vapour-phase collecting device has been completely flushed with ambient air and evacuated before the start of the smoking process. There shall not be any residual vacuum upstream of the collection device before smoking.

8.3.2 Smoking procedure

8.3.2.1 Smoke the cigarettes in accordance with the procedure stated in ISO 4387.

8.3.2.2 After completion of the smoking run, remove the cigarette butt and take five clearing puffs for each trap.

8.3.2.3 Record the total number of puffs taken on each channel, i.e. smoking puffs plus clearing puffs.

8.4 Measurement of carbon monoxide volume concentration

8.4.1 Recheck the calibration of the analyser (see 8.2.5) and introduce the vapour phase into the measuring cell of the analyser under the same conditions of ambient temperature and pressure as for sampling and the same gas flowrate as used during calibration. Read the analyser display which gives the carbon monoxide concentration.

8.4.2 At the end of each smoking, the vapour-phase collection container shall be emptied and flushed with air. The apparatus is then ready for the next smoking starting at step 8.3.2.1.

9 Expression of results

9.1 Calculation of the average carbon monoxide volume per cigarette

The average volume of carbon monoxide per cigarette is given by equation (1):

$$V_{as} = \frac{C \times V \times N \times p \times T_0}{S \times 100 \times p_0 \times (t + T_0)} \quad \dots (1)$$

where

- V_{as} is the average volume of carbon monoxide per cigarette, in millilitres;
- C is the percentage by volume of carbon monoxide observed;
- V is the puff volume, in millilitres;
- N is the number of puffs in the measured sample (including clearing puffs);
- p is the ambient pressure, in kilopascals;
- p_0 is the standard atmospheric pressure, in kilopascals;
- S is the number of cigarettes smoked;
- T_0 is the temperature for the triple point of water, in kelvin;
- t is the ambient temperature, in degrees Celsius.

For $V = 35$ ml and using rounded values of p_0 (101,3 kPa) and T_0 (273 K), equation (1) yields:

$$V_{as} = 0,9432 \times \frac{C \times N \times p}{S \times (t + 273)} \quad \dots (2)$$

9.2 Calculation of the average mass of carbon monoxide per cigarette

The average mass of carbon monoxide per cigarette is given by equation (3):

$$m_{cig} = \frac{C \times V \times N \times p \times T_0 \times M_{CO}}{S \times 100 \times p_0 \times (t + 273) \times V_m} \quad \dots (3)$$

where

- m_{cig} is the average mass of carbon monoxide per cigarette, in milligrams;
- M_{CO} is the molar mass of carbon monoxide, in grams per mole;
- V_m is the molar volume of an ideal gas, in litres per mole.

The other symbols are as defined in 9.1.

For $V = 35$ ml and using rounded values of M_{CO} (28 g/mol) and V_m (22,4 l/mol), equation (3) yields:

$$m_{cig} = 1,179 \times \frac{C \times N \times p}{S \times (t + 273)} \quad \dots (4)$$

10 Repeatability and reproducibility

An international collaborative study involving 32 laboratories and 4 samples, conducted in 1993, gave the following values for this method.

The difference between two single results found on matched cigarette samples by the same operator

using the same apparatus within the shortest feasible time interval will exceed the repeatability (r) on average not more than once in 20 cases in the normal and correct operation of the method.

Single results on matched cigarette samples reported by two laboratories will differ by more than the reproducibility (R) on average not more than once in 20 cases in the normal and correct operation of the method.

Data analysis gave the estimates as summarized in table 1.

Table 1

Mean yield of carbon monoxide mg	Repeatability r	Reproducibility R
3,45	0,47	1,18
3,56	0,42	1,03
9,89	0,85	2,22
13,80	1,09	3,00

For the purpose of calculating r and R , one test result was defined as the mean yield obtained from smoking 20 cigarettes in a single run.

NOTE 3 Further details are given in CORESTA Report 91/1, *The determination of repeatability and reproducibility for the measurement of NFDPM, nicotine, water using CORESTA methods 7, 8, 21, 22, 23 and 25.*

11 Test report

The test report shall show the method used and the results obtained. It shall also mention any operating conditions not specified in this International Standard or regarded as optional, as well as any circumstances that may have influenced the results.

The test report shall include all details required for complete identification of the sample.

If appropriate, the information listed in 11.1 to 11.4 shall be recorded.

11.1 Characteristic data about the cigarette and cigarette identification

- Name of manufacturer, country of manufacture.
- Product name.
- Date of sampling.
- Place of purchase or sampling.
- Kind of sampling point.
- Sampling point (e.g. address of retail outlet or machine number).

- Packet number (of that product sampled that day).
- Marks on any tax stamp.
- Printed smoke yields (if any).
- Length of cigarette.
- Length of filter.
- Length of overwrap.

11.2 Sampling

- Type of sampling procedure.
- Number of cigarettes in laboratory sample.
- Date and location of purchase.

11.3 Description of test

- Reference to this International Standard.
- Date of test.
- Type of smoking machine used.
- Type of analyser used.
- Total number of cigarettes smoked in the entire determination on that cigarette type.
- Number of cigarettes smoked into each collection device.
- Butt length.
- Room temperature (°C) during smoking operation and analysis.
- Relative humidity (%) during smoking operation.
- Atmospheric pressure (kPa) during smoking operation and analysis.

11.4 Test results

The expression of the laboratory data depends on the purpose for which the data are required, and the level of laboratory precision. Confidence limits shall be calculated and expressed on the basis of the laboratory data before any rounding has taken place.

- Average length of the cigarettes, to the nearest 0,1 mm.
- Average length of the filters, to the nearest 0,1 mm.
- Average length of the overwrap, to the nearest 0,1 mm.
- Average butt length to which the cigarettes were smoked, to the nearest 0,1 mm.
- Average length of tobacco portion smoked, to the nearest 0,1 mm.
- Average diameter of the cigarettes, in millimetres.
- Average draw resistance of the conditioned cigarettes.